

TSQH065



Transtick® used for the production of integral labels with a blood bag compliant adhesive. Key Selling Points 1. The adhesive is FDA compliant for blood bag labels. This product should be used when our label is the primary label affixed to the blood bag. Secondary labels do not require the same premium adhesive. 2. The form stock chosen should simulate a latex face stock. 3. New orders should be subject to test. A test should include the laser printing and the application of the label to the blood bag. 4. Please note that applying the label at room temperature is best. Any condensation on the bag could affect the adhesive performance.

- SL1055A 42# SCK Liner - Very Easy Release
- TC50614 Removable Adhesive Mid-Tack/ Blood Bag Compliant
- SL10040A 2.5 Mil Clay Coated Tight Release Laser Liner

Ship Date:	Min Order:	Master Width:	Liner Caliper:	Overall Caliper:	Shipping Weight:
1 week	1,000 msi ----- (645 m2)	25.5" (647mm) 51.5" (1308mm)	2.5 mil ----- (63 micron)	5.5 mil ----- (140 micron)	219 lbs/1000msi 154 kg/1000m2